



Slovak Society of Chemical Engineering  
Institute of Chemical and Environmental Engineering  
Slovak University of Technology in Bratislava

## PROCEEDINGS

51<sup>st</sup> International Conference of the Slovak Society of Chemical Engineering SSCHE 2025

Hotel DRUŽBA  
Jasná, Demänovská Dolina, Slovakia  
May 27 - 30, 2025

Editors: Assoc. Prof. Mário Mihaľ

ISBN: 978-80-8208-158-2, EAN: 9788082081582

Published by the Faculty of Chemical and Food Technology Slovak Technical University in Bratislava in Slovak Chemistry Library for the Institute of Chemical and Environmental Engineering; Radlinského 9, 812 37 Bratislava, 2024

Dudu, A., Covaci, E., Mot, A., Frentiu, T.: Limit of detection for Se, Zn, Cu, Cd, Pb, Hg and As determination by white small-sized electrothermal vaporization capacitively coupled microplasma optical emission spectrometry and sample combustion compared to inductively coupled plasma optical emission spectrometry, Editors: Mihaľ, M., In *51st International Conference of the Slovak Society of Chemical Engineering SSCHE 2025*, Jasná, Demänovská Dolina, Slovakia, 2025.

## **Limit of detection for Se, Zn, Cu, Cd, Pb, Hg and As determination by white small-sized electrothermal vaporization capacitively coupled microplasma optical emission spectrometry and sample combustion compared to inductively coupled plasma optical emission spectrometry**

<sup>1</sup>Adrian-Ioan Dudu\*, <sup>1,2</sup>Eniko Covaci, <sup>1</sup>Augustin-Catalin Mot, <sup>1,2</sup>Tiberiu Frentiu

<sup>1</sup>*Babes-Bolyai University, Faculty of Chemistry and Chemical Engineering, 11 Arany Janos, 400028, Cluj-Napoca, Romania*

<sup>2</sup>*Babes-Bolyai University, Research Centre for Advanced Analysis, Instrumentation and Chemometrics, Arany Janos 11, 400028 Cluj-Napoca, Romania*

*Corresponding author: e-mail: [adrian.dudu@ubbcluj.ro](mailto:adrian.dudu@ubbcluj.ro)*

**Key words:** capacitively coupled microplasma, inductively coupled plasma, optical emission spectrometry, sample combustion

The limit of detection (LODs) is considered the most important figure of merit for a method, that determines the suitability and applicability for various samples using different analytical systems. The aim of the current study was the LODs evaluation of a green and white analytical method based on sample combustion (50 mg sample combustion in O<sub>2</sub> atmosphere followed by uptake in 0.1 mol L<sup>-1</sup> HCl) and determination by small-sized electrothermal vaporization capacitively coupled microplasma optical emission spectrometry (SSETV- $\mu$ CCP-OES). A number of 100 episodic emission spectra of elements for an integration time of 100 ms were recorded using the High-Speed Acquisition mode of the Maya2000 Pro microspectrometer Spectra Suite software. The analytical signal was obtained by 2D signal integration (time-wavelength) for different pixel numbers around the Central Pixel (CP) of the analytical line (CP, CP $\pm$ 1, CP $\pm$ 2 and CP $\pm$ 3). It was found that the best LODs (mg kg<sup>-1</sup>) were obtained with CP $\pm$ 1 and CP $\pm$ 2, compared to CP strategy, namely 0.052 (Hg); 0.051 (Cu); 0.148 (Zn); 0.242 (Pb); 0.0042 (Cd); 0.322 (Se) and 0.173 (As). In terms of calibration curves linearity, it was found that the pixel numbers do not affect linearity, but affects the relative standard deviation of the blank (RSDB, %), becoming poorer as pixel numbers increases.

The LODs were also evaluated for inductively coupled plasma optical emission spectrometry (ICP-OES) with sample preparation based on classical microwave-assisted digestion in concentrated HNO<sub>3</sub> – H<sub>2</sub>O<sub>2</sub> mixture. Arsenic, Hg and Se were determined by ICP-OES after hydride generation with NaBH<sub>4</sub> solution. In the case of both methods, the LODs were also discussed in the light of current European Legislations for maximum admitted concentrations of toxic metals in foodstuffs and food supplements. It was found that, generally the SSETV- $\mu$ CCP-OES method is not suitable for toxic elements determination in vegetables and fruits, and it would necessitate its coupling with a preconcentration method, for example solid phase extraction (SPE) cartridges.

**Acknowledgements.** This work was supported by a grant of the Romanian Ministry of Research, Innovation and Digitization, CNCS/CCCDI-UEFISCDI, contract nr. 15PED/2025, project number PN-IV-P7-7.1-PED-2024-0091, within PNCDI IV.